	<b>Title</b>	<b>Instructions for Use Amorphous Hydrogel</b>
	<b>Revision</b>	B
	<b>Infocard Number</b>	<b>LFD-PKG-000248</b>

## Instructions for Use Amorphous Amorphous Hydrogel Wound Dressing

### 1.0 PRODUCT DESCRIPTION

Amorphous Hydrogel is a clear, viscous, sterile Amorphous Hydrogel containing a modified starch polymer, glycerol, preservatives and water.

Amorphous Hydrogel maintains a moist wound healing environment conducive to healing.

Dependant on wound conditions, Amorphous Hydrogel rehydrates dry necrotic tissue, to promote debridement in dry wounds, or absorbs wound slough and exudate where a small amount of exudate is present.

Amorphous Hydrogel can be washed away by irrigating the wound with sterile saline; this produces little or no trauma to the wound, and so results in improved wound healing and pain-free dressing changes.

### 2.0 INTENDED PURPOSE

The Amorphous Hydrogel can be used for prolonged use, as a non-invasive wound dressings intended principally for the management of most types of ulcers, pressure sores and other low exuding sloughy or necrotic wounds.

#### 2.1 INTENDED POPULATION

Individuals with the following wounds:

- Pressure ulcers
- Venous leg ulcers
- Diabetic foot ulcers
- First and second degree burns

#### 2.2 INTENDED USER

Intended for use by health professionals and may be used in a hospital

### 3. INDICATIONS

Amorphous Hydrogel is indicated for the management of dry non-exuding wounds and also low exuding, partial thickness wounds, such as

- Pressure ulcers
- Venous leg ulcers
- Diabetic foot ulcers
- First and second degree burns
- Incontinence Dermatitis

### 4. CONTRAINDICATIONS

- Ulcers resulting from infections, such as tuberculosis, syphilis, deep fungal Infections
- Third degree burns
- Heavily exuding wounds,
- Patients with known sensitivity to the Amorphous Hydrogel, or one of its components.
- babies under 12 weeks old.
- Do not use on areas where contact with the eyes is possible.


In case of infection with inflammatory signs (temperature, oedema, redness, pain), contact a doctor.


### 5. Warnings



Do not use if packaging is damaged or opened.

ARCHIVE

	<b>Title</b>	<b>Instructions for Use Amorphous Hydrogel</b>
	<b>Revision</b>	B
	<b>Infocard Number</b>	<b>LFD-PKG-000248</b>

 Do not re-use.

Re-use of single-use devices creates a potential risk to the patient or user. It may lead to contamination and/or impairment of functional capability. Contamination and/or limited functionality of the device may lead to injury, illness or death of the patient. Product is not suitable for re-sterilisation. For external use only and should not be used internally.

## 6.0 INFORMATION FOR USE

Amorphous Hydrogel is applied by first removing the lid. This may be a 'snap-off' or a screw lid, exposing the applicator nozzle then squeezing the tube and applying Amorphous Hydrogel directly to the wound area.

The interval between dressing changes will depend entirely upon the state of the wound. On exuding wounds, daily changes may be required at the beginning of treatment but this may be reduced to every 2 to 3 days for dry wounds after assessment by a suitably qualified healthcare professional.

### 6.1 Preparation

- a) Visually inspect the tube for damage prior to opening.
- b) Cleanse the wound with sterile saline or other suitable sterile cleaning solution with sterile swabs.
- c) Dry the skin surrounding the wound.

### 6.2 Amorphous Hydrogel Application

- a) Remove a sterile Amorphous Hydrogel tube from the unit box using local guidelines & best practices for handling sterile wound dressings. Inspect tube to ensure no damage or leaking Amorphous Hydrogel.
- b) Select the appropriate Amorphous Hydrogel size that will completely cover the wound surface, 'snap-off' the applicator nozzle cap or unscrew lid depending on product configuration.
- c) Wipe tip of nozzle with proprietary alcohol wipe.
- d) Keeping the nozzle clear of the wound surface, gently squeeze the tube, and apply Amorphous Hydrogel to the whole wound, to a depth of at least 5mm.
- e) Cover the wound with a suitable secondary dressing, compression therapy may be used in conjunction with Amorphous Hydrogel treatment, when so directed by a physician.
- f) Amorphous Hydrogel tubes are indicated for single use. Discard the tube with any unused Amorphous Hydrogel after first use.
- g) To minimize the risk of potential infection hazards or environmental pollution, disposable components of Amorphous Hydrogel should follow disposal procedures according to applicable and local laws, rules, regulations and infection prevention standards.

### 6.3 Dressing Change

Replace the secondary dressing if it becomes soiled, saturated or if exudate/drainage is observed. Otherwise, replace the dressing per established facility protocol.

- a) Gently remove Amorphous Hydrogel via irrigation using suitable sterile solution.
- b) Follow procedure 6.2 to apply a new dressing.

## 7. CLINICAL BENEFIT

The wound may initially appear to increase in size in the early stages of treatment with Amorphous Hydrogel wound dressing. This is normal and occurs as any wound debris is removed from the edges of the wound. This clears the way for healing.



**Title** Instructions for Use  
**Amorphous Hydrogel**  
**Revision** B  
**Infocard Number** LFD-PKG-000248

In the management of dry and low exuding wounds Amorphous Hydrogel can only make the overlying environment more conducive to healing. Instances where Amorphous Hydrogel treatment alone may make little or no progress, suitable treatment of the underlying conditions will be necessary as well. Therefore, if after 4-6 weeks of Amorphous Hydrogel treatment, there has been no improvement then, in line with accepted wound management practice, the original diagnosis and overall therapy should be reassessed with a healthcare professional.

## 8. STORAGE

Store dressings away from direct sunlight at ambient temperature and humidity.

## 9. DISPOSAL

To minimize the risk of potential infection hazards or environmental pollution, disposable components of Amorphous Hydrogel should follow disposal procedures according to applicable and local laws, rules, regulations and infection prevention standards.

## 10. SYMBOLS ON LABELLING

### WARNINGS



Do not use if the product sterile barrier system or its packaging is compromised



Medical Device



Single Sterile Barrier System

### STERILISATION



Method of sterilisation using irradiation



Do not re-use



Do not re-sterilise

### STORAGE



Keep away from direct sunlight

Keep dry

## 11. PRESENTATION

Amorphous Hydrogel is available in the following packaging configurations:

Description	Sizes	Units/Box	Article Number
Amorphous Hydrogel	8g	5	ML250P080S05
Amorphous Hydrogel	8g	10	ML250P080S10



**Title** **Instructions for Use**  
**Amorphous Hydrogel**  
**Revision** **B**  
**Infocard Number** **LFD-PKG-000248**

## 12. COMPLAINTS

For a patient/user/third party in the European Union and in countries with a similar regulatory system (Regulation (EU) 2017/745 on Medical Devices); if, during the use of this device or as a result of its use, a serious incident has occurred, please report it to the manufacturer and/or its authorized representative and to your national authority,

For complaints, questions or comments, contact Avery Dennison Medical Customer Support at phone +353 43 334 9586.



Manufactured By:  
Avery Dennison Medical Ltd.  
IDA Business Park, Ballinalee Road Longford, N39 DX73, Ireland

**Made in Ireland**



## 13. REVISION HISTORY

Revision	Change History	Originator	Date
A	Initial Revision, upload to master control	A Dennison	07/03/2022
B	Update format. Add sections 7,8 & 9	P Slattery	19/10/2022